

Certificate of Analysis

Product	BOULE CLEANING KIT (3 x 450 mL)
Product No	1504111
Lot No	2009-401
Manufacturing date	2020-07-13
Use before	2022-07-13

Test Hypochlorite	Specification
Hypochlorite concentration	2.0 – 2.4 %
Blank	PLT < 100 x 10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Test Enzymatic Cleaner	Specification
pH	7.6 – 8.4
Blank	PLT < 100 x 10 ⁹ /L RBC < 0.05 x 10 ¹² /L


Test Detergent Cleaner	Specification
pH	7.6 – 8.4
Blank	PLT < 100 x 10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Actual values meet specifications.

Date: 2020-09-08



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

